



Figure 4.17: Anomaly detection by means of an auto-encoder neural network. The mean absolute error is taken as the anomaly measure for the turbomolecular pump at position TT01 from March 1, 2018 until the failure on June 2, 2018. The green samples represent training samples, the blue dots are flagged as anomalies. The threshold is shown as an orange line.